



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/723,248	11/25/2003	Yves P. Arramon	PX-15	6049
21394 7590 01/26/2009 ARTHROCARE CORPORATION 7500 Rialto Boulevard Building Two, Suite 100 Austin, TX 78735-8532				
EXAMINER CUMBERLEDGE, JERRY L				
ART UNIT 3733		PAPER NUMBER		
NOTIFICATION DATE 01/26/2009		DELIVERY MODE ELECTRONIC		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

intel\_prop@arthrocure.com

# Office Action Summary

**Application No.**

10/723,248

**Applicant(s)**

ARRAMON, YVES P.

**Examiner**

JERRY CUMBERLEDGE

**Art Unit**

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 27 October 2008.  
2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.  
3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 25-34 is/are pending in the application.  
4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.  
5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.  
6) ☒ Claim(s) 25-34 is/are rejected.  
7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.  
8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.  
10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).  
11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
a) ☐ All b) ☐ Some \* c) ☐ None of:  
1. ☐ Certified copies of the priority documents have been received.  
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.  
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)  
2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)  
3) ☒ Information Disclosure Statement(s) (PTO/SF/US)  
Paper No(s)/Mail Date 10/27/2008  
4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_  
5) ☐ Notice of Informal Patent Application  
6) ☐ Other: \_\_\_\_\_

**DETAILED ACTION**

***Response to Amendment***

The declaration filed on 10/27/2008 under 37 CFR 1.131 is sufficient to overcome the Mazzuca et al. reference (US Pub. 2005/0070915 A1).

***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 25, 26, 32, 33 and 34 are rejected under 35 U.S.C. 102(b) as being anticipated by Chan (US Pat. 5,788,463).

Chan discloses an implant material injection system adapted for performing a percutaneous vertebroplasty procedure comprising: a remote actuator (Fig. 1, ref. 60); a pump (Fig. 1, ref. 40) comprising a piston (Fig. 1) and a drive chamber (Fig. 1, ref. 20), the pump having a distal end (Fig. 1) adapted to connect with a cannula, the drive chamber adapted to hold implant material (Fig. 1), the piston adapted to drive the implant material through the distal end of the drive chamber to an implant site (Fig. 1) (column 11, lines 29-54); a control line (Fig. 1, ref. 45) connecting the remote actuator and the pump, the control line adapted to advance the piston (Fig. 1); and wherein the implant material comprises a flowable hard tissue implant material (Fig. 2). The control line comprises a fluid column adapted to advance the piston (column 11, lines 49-54).

The system further comprising a cannula (Fig. 13, e.g. ref. 85) removably connected with the distal end of the drive chamber (Fig. 13-16). The system of further comprising an implant material reservoir connected with the pump (Fig. 1, ref. 30), the pump adapted to draw implant material from the material reservoir into the drive chamber (column 11, lines 29-54). The implant material comprises polymethylmethacrylate (Fig. 2)(column 11, lines 49-54).

With regard to statements of intended use and other functional statements, they do not impose any structural limitations on the claims distinguishable over the device of Chan, which is capable of being used as claimed if one so desires to do so. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963). Furthermore, the law of anticipation does not require that the reference "teach" what the subject patent teaches, but rather it is only necessary that the claims under attack "read on" something in the reference. *Kalman v. Kimberly Clark Corp.*, 218 USPQ 781 (CCPA 1983). Furthermore, the manner in which a device is intended to be employed does not differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. *Ex parte Masham*, 2 USPQ2d 1647 (1987). Specifically, the device is capable of moving fluids (Fig. 2)(column 11, lines 49-54) which would include a flowable hard tissue implant material (e.g. polymethylmethacrylate)

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and

the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 27-31 are rejected under 35 U.S.C. 103(a) as being unpatentable over Chan (US Pat. 5,788,463).

Chan discloses the claimed invention except for the control line has a length of about one foot. The control line has a length of about 36 inches. The control line has a length of at least 36 inches. The control line has a length of about 48 inches. The control line has a length greater than 48 inches. It would have been obvious to one having ordinary skill in the art at the time the invention was made to have constructed the control line at the above lengths, since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. In re Aller, 105 USPQ 233.

### ***Conclusion***

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Please see attached PTO-892.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to JERRY CUMBERLEDGE whose telephone number is (571)272-2289. The examiner can normally be reached on Monday - Friday, 8:30 AM - 5:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/J. C./  
Examiner, Art Unit 3733  
/Eduardo C. Robert/  
Supervisory Patent Examiner, Art Unit 3733